

Manual Supplement

Manual Title: 724 Users
Print Date: February 2000
Revision/Date: 1, 8/03

Supplement Issue: 7
Issue Date: 7/15
Page Count: 3

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 724
CD Rev. & Date: 2, 8/2003
CD PN: 1547849

Change #1

On page 41, add the following to the *Temperature, Thermocouples* table:

Type	Range	Measure and Source Accuracies (ITS-90)
XK	-200 to 800 °C	-200 to -100 ± 0.5 °C -100 to 800 ± 0.6 °C
BP	0 to 2500 °C	0 to 800 ± 1.2 °C 800 to 2500 ± 2.5 °C

Change #2

On page 16, under the *Note*, replace the first sentence with the following:

Available with V1.4 Firmware or greater.

Change #3, 41513

On page 43, under **General Specifications**, add the following:

Protection Class	Pollution Degree II
------------------	---------------------

Change #4

On page 37, Table 7:

Change:




18	Test lead, red	688051	1
	Test lead, black	688066	1

To:

18	Fluke-7XX Test Lead Set	3397308	1
----	-------------------------	---------	---

Change #5, 67401, 169, 512

On page 7, under **Table 2**, add:

Symbol	Meaning
	Consult user documentation.
	Conforms to relevant Australian EMC standards.
	Conforms to relevant South Korean EMC Standards.

On page 43, under **General Specifications**, replace the Safety section and add:

Safety	IEC 61010-1: Pollution Degree 2
Electromagnetic Compatibility (EMC)	
International	<p>IEC 61326-1: Portable Electromagnetic Environment CISPR 11: Group 1, Class A</p> <p><i>Group 1: Equipment has intentionally generated and/or uses conductively-coupled radio frequency energy that is necessary for the internal function of the equipment itself.</i></p> <p><i>Class A: Equipment is suitable for use in all establishments other than domestic and those directly connected to a low-voltage power supply network that supplies buildings used for domestic purposes. There may be potential difficulties in ensuring electromagnetic compatibility in other environments due to conducted and radiated disturbances.</i></p> <p><i>Emissions that exceed the levels required by CISPR 11 can occur when the equipment is connected to a test object.</i></p>
Korea (KCC)	<p>Class A Equipment (Industrial Broadcasting & Communication Equipment)</p> <p><i>Class A: Equipment meets requirements for industrial electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and not to be used in homes.</i></p>
USA (FCC)	47 CFR 15 subpart B. This product is considered an exempt device per clause 15.103.